

MEETING ABSTRACT

Open Access

The performance of silicon photomultipliers in Cherenkov TOF PET

Rok Dolenc^{1*}, Samo Korpar¹, Peter Krizan², Rok Pestotink²

From PSMR 2015: 4th Conference on PET/MR and SPECT/MR
La Biodola, Isola d'Elba, Italy. 17-21 May 2015

¹Faculty of Chemistry and Chemical Engineering, University of Maribor, Maribor, Slovenia

In time-of-flight positron emission tomography (TOF PET) one of the main factors limiting the time resolution is the time evolution of the scintillation process. This can be avoided by using exclusively the Cherenkov light produced in a suitable material. Sub 100 ps FWHM timing has already been experimentally demonstrated but with a drawback of relatively low detection efficiency due to the photodetectors used. In this work silicon photomultipliers (SiPMs) are considered as a photodetector in Cherenkov TOF PET. The detection efficiency can be significantly improved by using SiPMs, however, at room temperature the SiPM dark counts introduce a significant source of fake coincidences. SiPM samples from different producers were tested in a simple back-to-back setup in combination with lead fluoride Cherenkov radiators. Results for coincidence timing, detection efficiency and effects of dark counts at different temperatures and SiPM overvoltages are presented.

Authors' details

¹Faculty of Chemistry and Chemical Engineering, University of Maribor, Maribor, Slovenia. ²Jozef Stefan Institute, Ljubljana, Slovenia.

Published: 18 May 2015

doi:10.1186/2197-7364-2-S1-A1

Cite this article as: Dolenc et al.: The performance of silicon photomultipliers in Cherenkov TOF PET. *EJNMMI Physics* 2015 **2**(Suppl 1):A1.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Immediate publication on acceptance
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► springeropen.com