

MEETING ABSTRACT

Open Access

Iterative reconstruction incorporating positron range correction within STIR framework

Ottavia Bertolli^{1,2*}, Matteo Cecchetti², Niccolò Camarlinghi², Afroditi Eleftheriou^{1,3}, Nicola Belcari², Charalampos Tsoumpas¹

From PSMR14: 3rd Conference in PET/MR and SPECT/MR
Kos Island, Greece. 19-21 May 2014

¹Division of Medical Physics,
University of Leeds, UK

Positron range is one important physical limitation to spatial resolution in PET imaging. We present a method to take account of positron range blurring in iterative reconstruction by including specific positron range kernels approximating the annihilation distribution in the forward projection operation [1]. The correction method is implemented within the OSMAPOSL iterative reconstruction algorithm of STIR 3.0 [2].

GATE [3] simulation package was used to obtain the positron annihilation distribution of a point source of ⁶⁸Ga placed in the centre of spherical water and lung tissue phantoms. The simulations were performed without and with 3T static magnetic field in the axial direction. To obtain the kernels the annihilation coordinates were voxelized into a 3D matrix with element size equal to the reconstructed image's voxel size and then normalised. The proposed method was evaluated using simulation data (generated with STIR) of a parallelepiped water phantom containing two spherical hot spots, on the HYPER-Image preclinical PET/MR scanner [4]. To simulate the positron range effect the phantom was blurred by convolution with the positron kernel relative to water. The blurred images were reconstructed with and without filtering of the forward projected image.

Line profiles traced through the spots of the reconstructed images show that the use of the positron range correction yields sharper boundary definition, albeit with noise enhancement. The values relative to the spots volume show that the correction results in the recovery of the activity mean value compared to the simulated phantom (e.g. 55% to 94% for the smaller spot, no magnetic field).

Furthermore the tissue-dependent correction, which consists in blurring the image with the kernel correspondent to the voxel tissue (water or lung), was implemented and evaluated on a phantom with water and lung regions. The activity mean value is recovered but the method performance in correspondence to the edges needs further investigation.

Authors' details

¹Division of Medical Physics, University of Leeds, UK. ²Department of Physics, University of Pisa and INFN Pisa, Italy. ³Department of Physics, National and Kapodistrian University of Athens, Greece.

Published: 29 July 2014

References

1. Cal-Gonzalez J, Herraiz JL, et al: Positron range effects in high resolution 3D PET imaging. *IEEE Nucl. Sci. Symp. Med. Imag. Conf. Rec* 2009, 2788-91.

2. Thielemans K, et al: **STIR: software for tomographic image reconstruction release 2.** *Phys Med Biol* 2012, **57**(4):867-883.
3. Jan S, et al: **GATE: a simulation toolkit for PET and SPECT.** *Phys Med Biol* 2004, **49**:4543-4561.
4. Schulz V, et al: **A preclinical PET/MR insert for a human 3T MR scanner.** *IEEE Nucl. Sci. Symp. Med. Imag. Conf. Rec.* 2009, **1-5**:2577-2579.

doi:10.1186/2197-7364-1-S1-A42

Cite this article as: Bertolli et al.: Iterative reconstruction incorporating positron range correction within STIR framework. *EJNMMI Physics* 2014 **1**(Suppl 1):A42.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- ▶ Convenient online submission
- ▶ Rigorous peer review
- ▶ Immediate publication on acceptance
- ▶ Open access: articles freely available online
- ▶ High visibility within the field
- ▶ Retaining the copyright to your article

Submit your next manuscript at ▶ springeropen.com
