

MEETING ABSTRACT

Open Access

Dynamic brain PET/MR using TOF reconstruction

Mohammad Mehdi Khalighi^{1*}, Gaspar Delso², Michel Tohme¹, Andrei Iagaru³, Greg Zaharchuk¹

From PSMR 2015: 4th Conference on PET/MR and SPECT/MR
La Biodola, Isola d'Elba, Italy. 17-21 May 2015

¹GE Healthcare, USA

In a functional PET/MR study, it is difficult to get good temporal resolution of activity distribution from PET images because of the need to image for a certain length of time to get sufficient count statistics (image SNR). Time-of-flight (TOF) reconstruction can be used to increase PET images SNR and therefore increase the temporal resolution. Five patients were injected with 410 ± 80 MBq of FDG and scanned 140 ± 30 minutes post-injection on a simultaneous TOF-enabled PET/MR scanner. PET images were reconstructed with and without TOF. TOF reconstruction shows faster convergence while it achieves a temporal SNR improvement of 5-45% ($25 \pm 5\%$) compared to non-TOF reconstruction. With this additional SNR gain, frame durations as short as 30s are possible while preserving reasonable image quality. This in turn effectively increases the temporal resolution of dynamic brain studies using simultaneous PET/MR imaging.

Authors' details

¹GE Healthcare, USA. ²UniversitätsSpital Zürich, Switzerland. ³Stanford University, USA.

Published: 18 May 2015

doi:10.1186/2197-7364-2-S1-A60

Cite this article as: Khalighi et al.: Dynamic brain PET/MR using TOF reconstruction. *EJNMMI Physics* 2015 **2**(Suppl 1):A60.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Immediate publication on acceptance
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► springeropen.com